

## ELECTRICAL LISTING PROGRAM

Customer: Bonal Technologies, Inc.  
 Class: Laboratory Equipment  
 Location: Royal Oak, MI, USA  
 Listing No. E11141  
 Report No. E11141-2001 Ed. 3  
 Project No. E11141-2102  
 Effective Date: November 17, 2020  
 Last Revised Date: October 21, 2021  
 Expires: N/A

Standards: CAN/CSA-C22.2 No. 61010-1 “Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General requirements”

UL 61010-1 “Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General requirements”

Product: Vibration Stress Relief and Weld Conditioning System

Markings: Products are marked in a permanent manner where it is readily visible after installation with the following:

1. Manufacturer’s name, trademark or other recognized symbol of identification.
2. Model number or designation.
3. Electrical ratings: voltage in V, frequency in Hz, input in VA.
4. Month and year of manufacture. Date coding, serial number or equivalent means may be used.
5. QAI Listing Mark:



6. QAI Listing Number: E11141.
7. The statement “Conforms to CSA 61010-1 and UL 61010-1”.

Models / Ratings:

<b>Series: Meta-Lax</b>	
<b>Model</b>	<b>Rating</b>
2800&2800S	110 Vac, 50/60 Hz, 1650 VA/1200 VA

The materials, products or systems listed herein have been qualified to bear the QAI Listing Mark under the conditions stated with each Listing. Only those products bearing the QAI Listing Mark are considered to be listed by QAI. No warrantee is expressed or implied, and no guarantee is provided that any jurisdictional authority will accept the Listing found herein. The appropriate authorities should be contacted regarding the acceptability of any given Listing. Visit the QAI Online Listing Directory located at [www.qai.org](http://www.qai.org) for the most up to date version of this Listing and to validate that this QAI Listing is active. Questions regarding this listing may be directed to [info@qai.org](mailto:info@qai.org). Please include the listing number in the request.

\*\*\*