

ELECTRICAL LISTING PROGRAM

Customer: Phase Analyzer Company
Class: Laboratory Equipment
Location: Richmond, BC, Canada
Website: <http://www.phase-technology.com/>
Listing No. E11098
Report No. E11098-1901
Project No. E11098-1901
Effective Date: January 20, 2020
Last Revised N/A
Date:
Expires: N/A

Standards: CSA C22.2 No. 61010-1 & UL 61010-1 "Safety requirements for Electrical Equipment for Measurement, Control, and Laboratory Use"

Product: Laboratory Equipment use as Petroleum Analyzer to measure viscosity and density.

Markings: Products are marked in a permanent manner where it is readily visible after installation with the following:

- Manufacturer's name or trademark
- Model designation
- Month and year of manufacture
- Product ratings: voltage range, frequency and current
- QAI file Number: E11098
- The wording "CSA C22.2 No. 61010-1 and UL 61010-1"
- QAI logo shown here:



Models / Ratings:

Series: N/A			
Model	Rating	Description	Environmental Locations
OptiMVD-AS1	100-240V, 50-60Hz, 2.4-5A	Single Carousel and Single Heated Enclosure	Dry Location only
OptiMVD-AS2	100-240V, 50-60Hz, 2.4-5A	Dual Carousel and Dual Heated Enclosure	Dry Location only

Notes: N/A

The materials, products or systems listed herein have been qualified to bear the QAI Listing Mark under the conditions stated with each Listing. Only those products bearing the QAI Listing Mark are considered to be listed by QAI. No warrantee is expressed or implied, and no guarantee is provided that any jurisdictional authority will accept the Listing found herein. The appropriate authorities should be contacted regarding the acceptability of any given Listing. Visit the QAI Online Listing Directory located at www.qai.org for the most up to date version of this Listing and to validate that this QAI Listing is active. Questions regarding this listing may be directed to info@qai.org. Please include the listing number in the request.
