

## ELECTRICAL LISTING PROGRAM

Customer: Longevity Resources Inc.  
 Class: Ozone Generator  
 Location: North Saanich, British Columbia, Canada  
 Listing No.: E10538  
 Report No.: E10538-1401 Edition 6  
 Project No.: E10538-2101  
 Effective Date: October 22, 2014  
 Last Revised Date: March 1, 2021  
 Expires: N/A

Standards: CSA C22.2 No. 61010-1-12 3<sup>rd</sup> Edition - Safety Requirements for Electrical Equipment for measurement, Control, and Laboratory Use – Part 1: General Requirements.  
 UL 61010-1 3<sup>rd</sup> Edition - Safety Requirements for Electrical Equipment for measurement, Control, and Laboratory Use – Part 1: General Requirements.

Product: Ozone Generator  
 Cord-Connected, Dry Locations and Laboratory use only.

Markings: Products are marked in a permanent manner where it is readily visible after installation with the following:

- a) Manufacturer’s name or trademark
- b) Model designation
- c) Month and year of manufacture
- d) Product ratings: voltage, frequency, ampere
- e) QAI file Number: E10538
- f) The wording “Conforms to CSA C22.2 No. 61010-1 and UL 61010-1”
- g) Warning markings as required by the standards in English and French

h) QAI Logo 

Models / Ratings:

Model	Volts VAC	Current A	Volts-Ampere VA	Hz	Location
QUANTUM 5 (MOU120)	120	0.9	108	60	Dry Location
QUANTUM 3 (DOU120)	120	0.9	108	60	Dry Location
HTU-500G (EXT120T)	120	0.9	108	60	Dry Location
HTU-500G2 (EXT120)	120	0.9	108	60	Dry Location
QOG (EXT120 Ultra)	120	0.3	36	60	Dry Location
QOG-T (EXT120-T Ultra)	120	0.3	36	60	Dry Location
EXT50	100/230	0.3	30/69	50/60	Dry Location
Pulsar	100/230	0.3	30/69	50/60	Dry Location



# LABORATORIES

CERTIFICATION TESTING INSPECTION

VANCOUVER, BC: 877.461.8378 ph. | 604.527.8368 fx.  
 LOS ANGELES, CA: 909.483.0250 ph. | 909.483.0336 fx.  
 WASHINGTON, DC: 540.636.9445 ph. | 540.636.9414 fx.  
 TULSA, OK: 918.437.8333 ph. | 918.437.8487 fx.  
 TORONTO, ON: 905-605-5444  
 SEATTLE, WA: 425.512.8419  
 WEBSITE: WWW.QAI.ORG

QUANTUM 3 "EPIC" (DOU "EPIC")	100/230	0.9	90/207	50/60	Dry Location
QUANTUM 5 "EPIC" (MOU "EPIC")	100/230	0.9	90/207	50/60	Dry Location

Note: 1- Product is intended for laboratory use only.

2- This product was evaluated for electrical safety only; the physiological effect of ozone was not considered.

The materials, products or systems listed herein have been qualified to bear the QAI Listing Mark under the conditions stated with each Listing. Only those products bearing the QAI Listing Mark are considered to be listed by QAI. No warrantee is expressed or implied, and no guarantee is provided that any jurisdictional authority will accept the Listing found herein. The appropriate authorities should be contacted regarding the acceptability of any given Listing. Visit the QAI Online Listing Directory located at [www.qai.org](http://www.qai.org) for the most up to date version of this Listing and to validate that this QAI Listing is active. Questions regarding this listing may be directed to [info@qai.org](mailto:info@qai.org). Please include the listing number in the request.

\*\*\*