



## INSPECTION BODY ACCREDITATION PROGRAM (IBAP)

### Scope of Accreditation

**Accredited Legal Entity:** QAI LABORATORIES LTD.

Contact Name: Lawrence Gibson

Address: 3980 North Fraser Way  
Burnaby, BC V5J 5K5

Telephone: +1 604 527 8378

Fax: +1 604 527 8368

Website: <http://www.qai.org>

Email: [lqibson@qai.org](mailto:lqibson@qai.org)

<b>Accreditation Number:</b>	06008
<b>Initial Accreditation Date:</b>	2010-01-15
<b>Reaccreditation Date:</b>	2014-01-21
<b>Accreditation Expiry Date:</b>	2018-01-15
<b>Accreditation Standards:</b>	ISO/IEC 17020:2012 Requirements and Guidance - Inspection Body Accreditation Program

### SPE-1000 Field Evaluation/Special Inspection

#### Inspection Label:





**Scope of Accreditation:**

The performing of field inspections and the application of approval labels on electrical equipment and systems in accordance with all of the following:


As identified in the scope of CSA SPE-1000 Model Code for the field evaluation of electrical equipment, but excluding

- a) wire and cable products;
- b) wiring devices;
- c) equipment for use in hazardous locations;
- d) medical electrical equipment and systems;
- e) components\* that will require further evaluation as part of a complete assembly, such as switches, relays, and timers;
- f) equipment connected to line voltage in excess of 46 kV;
- g) manlifts, elevators, climb assists and similar systems (other than their associated control panels); and
- h) any equipment that is not permitted to be field evaluated as directed by an AHJ (such as air-cleaning equipment that intentionally produces ozone).

\*Special inspection/field evaluation only applies to equipment that is capable of being connected to the supply in accordance with the Canadian Electrical Code, Part I.

**SPE-3000 Field Evaluation/Special Inspection**

**Inspection label:**



**Scope of Accreditation:**

The performing of field evaluations/special inspections and the application of approval labels on medical-electrical equipment (MEE) and medical-electrical systems (MES)+ in accordance with all of the following:



As identified in the scope of CSA SPE-3000 Model Code for the field evaluation/special inspection of medical electrical equipment and systems, but excluding the following product categories at this time:

- a) cosmetic and hygiene equipment;
- b) X-ray equipment and systems;
- c) MRIs and CT scans; and
- d) laser equipment.

\*The definitions for Medical electrical equipment (MEE) and Medical electrical system (MES) can be found in SPE-3000-15 Section 3 Definitions.