

## ELECTRICAL LISTING PROGRAM

Customer: Jelight Company, Inc.  
 Class: Laboratory  
 Location: Irvine, CA  
 Website: <http://www.jelight.com/>  
 Listing No. E10956  
 Report No. E10956-1801 Ed. 2  
 Project No. E10956-1802  
 Effective Date: May 10th, 2018  
 Last Revised Date: October 27<sup>th</sup>, 2018  
 Expires: N/A

Standards: UL 61010-1 - Third Edition: Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements – 2012/05/11  
 CSA-C22.2 NO. 61010-1-12 (R2017) - Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements

Product: The products covered by this report are hand-held, portable UV curing systems. The curing system consists of a power unit, ultraviolet (UV) light irradiator and docking station. The irradiator provides a UV light source and the docking station allows temporary placement of the Irradiator without having to extinguish the lamp.

Markings: Products are marked in a permanent manner where it is readily visible after installation with the following:

- a) Manufacturer's name or trademark
- b) Model designation
- c) Month and year of manufacture
- d) Product ratings: voltage, frequency, current
- e) QAI file Number: E10956
- f) The wording "UL 61010-1 and CSA C22.2 NO. 61010-1"
- g) Warning markings as required by the standards in English and French
- h) QAI logo shown here:



Models / Ratings:

Model	Rating	Description	Environmental Locations
JUVC-5B	120VAC, 60Hz, 15A	Portable UV Curing Equipment	Dry Location only
JUVC-5BX	120VAC, 60Hz, 15A	Portable UV Curing Equipment	Dry Location only