

VANCOUVER, BC: LOS ANGELES, CA: TULSA, OK: TORONTO, ON WEBSITE: 877.461.8378 ph. | 604.527.8368 fx. 909.483.0250 ph. | 909.483.0336 fx. 918.437.8333 ph. | 918.437.8487 fx. 905.605.5444 WWW.QAL.ORG

ELECTRICAL LISTING PROGRAM

Customer: AZCO Industries Ltd. Class: Laboratory Equipment Location: Langley, BC, Canada

Listing No. E10552 Report No. E10552-2101 Project No. E10552-2101 Effective Date: 2019-01-07 Last Revised Date: 2022-07-14 Expires: N/A

> Standards: CAN/CSA C22.2 No. 61010-1 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements.

> > UL 61010-1 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements.

- Product: Ozone Generator; Cord-Connected, Laboratory, Dry Locations Only
- Markings: Products are marked in a permanent manner where it is readily visible after installation with the following:
 - a) Manufacturer's name or trademark
 - b) Model designation
 - c) Month and year of manufacture, date code, serial number, or equivalent method of unique identification
 - d) Product ratings: voltage, frequency, amperes
 - e) QAI file Number: E10552
 - f) The wording "Conforms to CAN/CSA-C22.2 No. 61010-1 and UL 61010-1"
 - g) Warning markings as required by the standards in English and French



h) QAI logo shown here: ^c

Models / Ratings:	Model #	Volts	Current	Hz	Location
	HTU-500	120	0.4	50/60	Indoor Dry



VANCOUVER, BC: 877.461.8378 ph. | 604.527.8368 fx. LOS ANGELES, CA: 909.483.0250 ph. | 909.483.0336 fx. TULSA, OK: 918.437.8333 ph. | 918.437.8487 fx. TORONTO, ON 905.605.5444 WEBSITE: WWW.QAI.ORG

- Notes: 1. The manufacturer / Listee maintains a system to ensure each product complies with this report prior to application of the QAI mark.
 - 2. The product is installed in accordance with the NEC and CEC as applicable.
 - 3. Suitability of the product in its final installation is subject to inspection by the Authority Having Jurisdiction (AHJ).
 - 4. This product was evaluated for electrical safety only; the physiological effect of ozone was not considered.
 - 5. Product is intended for laboratory use only.

The materials, products or systems listed herein have been qualified to bear the QAI Listing Mark under the conditions stated with each Listing. Only those products bearing the QAI Listing Mark are considered to be listed by QAI. No warrantee is expressed or implied, and no guarantee is provided that any jurisdictional authority will accept the Listing found herein. The appropriate authorities should be contacted regarding the acceptability of any given Listing. Visit the QAI Online Listing Directory located at <u>www.qai.org</u> for the most up to date version of this Listing and to validate that this QAI Listing is active. Questions regarding this listing may be directed to <u>info@qai.org</u>. Please include the listing number in the request.
