



ELECTRICAL LISTING PROGRAM

Customer: AZCO Industries Ltd.
Class: Laboratory Equipment
Location: Langley, BC, Canada

Listing No. E10552
Report No. E10552-2301
Project No. E10552-2301
Effective Date: 2023-06-30
Last Revised N/A
Expires: N/A

Standards: CAN/CSA C22.2 No. 61010-1-12 Ed.3 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements.

UL 61010-1 Ed.3 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements.

Product: Air Dryer, Cord-Connected, Laboratory, Dry Locations Only

Markings: Products are marked in a permanent manner where it is readily visible after installation with the following:

- a) Manufacturer’s name or trademark
- b) Model designation
- c) Month and year of manufacture
- d) Product ratings: voltage, frequency, watts
- e) QAI file Number: E10552
- f) The wording “CSA CAN/CSA-C22.2 No. 61010-1-12 Ed.3 and UL 61010-1 Ed.3”
- g) Warning markings as required by the standards in English and French
- h) QAI logo shown here:



Models / Ratings:

Model	Rating	Description	Environmental Locations
VMD-12	120 ,60Hz /240, 50Hz	Max 2A	Indoor & Dry
VMD-32	120 ,60Hz /240, 50Hz	Max 3A	Indoor & Dry