



ELECTRICAL LISTING PROGRAM

Customer: 13983129 Canada Inc (Aquassist)
Class: Laboratory Instruments
Location: Boucherville, QC, Canada
Listing No. E11290
Report No. E11290-2201
Project No. E11290-2201
Effective Date: June 20, 2023
Last Revised N/A
Date:
Expires: N/A

Standards: **For PAK:**

CSA C22.2 No.62368-1:19 3rd Edition –Audio/Video, Information and Communication Technology Equipment – Safety – Part 1: General Requirements

UL 62368-1 3rd Edition – Audio/Video, Information and Communication Technology Equipment – Safety – Part 1: General Requirements

For PSA, PTA, PEA-P, PEA-C:

UL 61010-1 Third Edition & CSA C22.2 No.61010-1-12 Third Edition – Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use

With reference to:

UL 61010-2-051 & CSA C22.2 No 61010-2-051 – Laboratory Equipment for mixing and stirring

UL 61010-2-081 & CSA C22.2 No 61010-2-081 – Automatic and Semi-Automatic laboratory equipment for analysis and other purposes

Product: Electrical Equipment for Measurement, Control and Laboratory Use
Markings: Products are marked in a permanent manner where it is readily visible after following installation with the:

- a) Manufacturer's name or trademark
- b) Model designation
- c) Month and year of manufacture
- d) Product ratings: voltage, frequency, watts
- e) QAI file Number: E11290
- f) The wording **For PAK:** "Conforms to/Conformé a: CSA C22.2 No. 62368-1: Ed. 3 and UL 62368-1: Ed. 3"
For PTA: "Conforms/Conformé a: CSA C22.2 No. 61010-1 and UL 61010-1, CSA C22.2 No 61010-2-051 & UL 61010-2-051 and CSA C22.2 No 61010-2-081 & UL 61010-2-081"
For PSA, PEA-P, PEA-C: "Conforms/Conformé a: CSA C22.2 No. 61010-1 and UL 61010-1 and CSA C22.2 No 61010-2-081 & UL 61010-2-081" "CSA C22.2 No. 61010-1 and UL 61010-1"
- g) Warning markings as required by the standards in English and French>



h) QAI logo shown here:



Models / Ratings:

Model	Rating	Description	Environmental Locations
PAK	100-240 VAC 1.2A 80W	Portable Analyzer Kit	Indoor
PTA	3.7VDC, 2600mAh	Portable Titration Analyzer	Indoor
PSA	3.7VDC, 2600mAh	Portable Spectrometry Analyzer	Indoor
PEA-P	3.7VDC, 850mAh	Portable pH Analyzer	Outdoor IP53
PEA-C	3.7VDC, 850mAh	Portable Conductivity Analyzer	Outdoor IP53

The materials, products or systems listed herein have been qualified to bear the QAI Listing Mark under the conditions stated with each Listing. Only those products bearing the QAI Listing Mark are considered to be listed by QAI. No warranty is expressed or implied, and no guarantee is provided that any jurisdictional authority will accept the Listing found herein. The appropriate authorities should be contacted regarding the acceptability of any given Listing. Visit the QAI Online Listing Directory located at www.qai.org for the most up to date version of this Listing and to validate that this QAI Listing is active. Questions regarding this listing may be directed to info@qai.org. Please include the listing number in the request.