



ELECTRICAL LISTING PROGRAM

Customer: Yourgene Health UK Ltd.
Class: Laboratory Equipment
Location: Manchester, UK
Website: <https://www.yourgene-health.com>

Listing No. E11269
Report No. E11269-2205 Ed.2
Project No. E11269-2404
Effective Date:
Last Revised November 20th, 2024
Date:
Expires: N/A

Standards: CSA C22.2 No. 61010-1 “Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements”

UL 61010-1 “Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements”

Product: Laboratory Equipment

Markings: Products are marked in a permanent manner where it is readily visible after installation with the following:

- a) Manufacturer’s name or trademark
- b) Model designation
- c) Month and year of manufacture
- d) Product ratings: voltage, frequency, watts
- e) QAI file Number: E10567
- f) The wording “Conforms to CSA C22.2 No. 61010-1 and UL 61010-1”
- g) Warning markings as required by the standards in English and French>
- h) QAI logo shown here:





Models / Ratings:

Series: N/A			
Model	Rating	Description	Environmental Locations
NIMBUS (PN CG- 13900-xx)	110- 240VAC, 5.0A, 50/60Hz	Laboratory Equipment, automate agarose gel DNA electrophoresis	Indoor

Notes: xx is 01 to 12

The materials, products or systems listed herein have been qualified to bear the QAI Listing Mark under the conditions stated with each Listing. Only those products bearing the QAI Listing Mark are considered to be listed by QAI. No warrantee is expressed or implied, and no guarantee is provided that any jurisdictional authority will accept the Listing found herein. The appropriate authorities should be contacted regarding the acceptability of any given Listing. Visit the QAI Online Listing Directory located at www.qai.org for the most up to date version of this Listing and to validate that this QAI Listing is active. Questions regarding this listing may be directed to info@qai.org. Please include the listing number in the request.
