



## ELECTRICAL LISTING PROGRAM

Customer: The Magstim Company Limited  
Class: Medical Equipment  
Location: Spring Gardens  
Whitland, Carmarthenshire  
United Kingdom, SA34 0HR

Listing No. E11381  
Report No. E11381-2401  
Project No. E11381-2401  
Effective Date: Oct 10<sup>th</sup>, 2025  
Last Revised Date: Oct 10<sup>th</sup>, 2025  
Expires: N/A

Standards: IEC 60601-1:2005/AMD2:2020 – Medical electrical equipment – Part 1:  
General requirements for basic safety and essential performance

CAN/CSA C22.2 No. 60601-1:14 – Medical electrical equipment – Part 1:  
General requirements for basic safety and essential performance

AAMI ES60601-1:2005/AMD2:2021 – Medical electrical equipment – Part  
1: General requirements for basic safety and essential performance

Product: Rapid Systems – Nerve stimulation system

Markings: Products are marked in a permanent manner where it is readily visible  
after installation with the following:

- a) Manufacturer's name or trademark
- b) Model designation
- c) Month and year of manufacture
- d) Product ratings: voltage, frequency, watts
- e) QAI file Number: E11381
- f) Warning markings as required by the standards in English and French
- g) QAI logo shown here:





Models /  
Ratings:

| <b>Horizon Systems</b>  |                         |  |                                |
|-------------------------|-------------------------|--|--------------------------------|
| <b>Model</b>            | <b>Rating</b>           | <b>Description</b>   | <b>Environmental Locations</b> |
|                         | 230/240Vac,<br>50/60Hz, | A medical device that stimulates cortical and peripheral nerves for therapeutic, diagnostic and research purposes. | IPXX                           |
| Super Rapid Plus System | 115Vac,<br>60Hz         |  |                                |
|                         |                         |  |                                |
|                         |                         |  |                                |
|                         |                         |  |                                |
|                         |                         |  |                                |

The materials, products or systems listed herein have been qualified to bear the QAI Listing Mark under the conditions stated with each Listing. Only those products bearing the QAI Listing Mark are considered to be listed by QAI. No warrantee is expressed or implied, and no guarantee is provided that any jurisdictional authority will accept the Listing found herein. The appropriate authorities should be contacted regarding the acceptability of any given Listing. Visit the QAI Online Listing Directory located at [www.qai.org](http://www.qai.org) for the most up to date version of this Listing and to validate that this QAI Listing is active. Questions regarding this listing may be directed to [info@qai.org](mailto:info@qai.org). Please include the listing number in the request.