

VANCOUVER, BC: LOS ANGELES, CA: TULSA, OK: TORONTO, ON WEBSITE:

877.461.8378 ph. | 604.527.8368 fx. 909.483.0250 ph. | 909.483.0336 fx. 918.437.8333 ph. | 918.437.8487 fx. 905.605.5444 WWW.OALORG

## **ELECTRICAL LISTING PROGRAM**

Customer: The Magstim Company Limited Class: Medical Equipment Location: Spring Gardens whiteland, Carmarthenshire, UK SA34 0HR Website: www.Magstim.com

Listing No. E11381 Report No. E11381-2401 Project No. E11381-2401 Effective February 24<sup>th</sup>, 2025 Date: Last Revised February 24<sup>th</sup>, 2025 Date: Expires: N/A

Standards: IEC 60601-:2005/AMD2:2020 - Medical electrical equipment – Part 1: General requirements for basic safety and essential performance

CAN/CSA-C22.2 No. 60601-1:14 – Medical electrical equipment – Part 1: General requirements for basic safety and essential performance

AAMI ES60601-1:2005/AMD2:2021 - Medical electrical equipment – Part 1: General requirements for basic safety and essential performance

- Product: Rapid Systems Nerve stimulation system
- Markings: Products are marked in a permanent manner where it is readily visible after installation with the following:
  - a) Manufacturer's name or trademark
  - b) Model designation
  - c) Month and year of manufacture
  - d) Product ratings: voltage, nature of supply, frequency, and amps
  - e) QAI file Number: E11381
  - f) Warning markings as required by the standards in English and French
  - g) QAI logo shown here:





VANCOUVER, BC: 877.461.8378 ph. | 604.527.8368 fx. LOS ANGELES, CA: 909.483.0250 ph. | 909.483.0336 fx. TULSA, OK: 918.437.8333 ph. | 918.437.8487 fx. TORONTO, ON 905.605.5444 WEBSITE: WWW.QALORG

	Model #	Rated Input V	Rated Frequency	Wattage or Current	Description	Environmental Condition
1.			50Hz	N/A	Medical	IPXX
			60Hz		device that	
2.	Super Rapid <sup>2</sup>				delivers	
3.	Super Rapid <sup>2</sup>				magnetic	
	Plus 1				stimulation.	
4.	Rapid <sup>2</sup>					
	Therapy					
	System					
5.	Magstim					
	Standard					
	Rapid <sup>2</sup>					
6.	Magstim					
	Super Rapid <sup>2</sup>					
7.	Magstim					
	Rapid <sup>2</sup>					
	Plus1System					

h) The materials, products or systems listed herein have been qualified to bear the QAI Listing Mark under the conditions stated with each Listing. Only those products bearing the QAI Listing Mark are considered to be listed by QAI. No warrantee is expressed or implied, and no guarantee is provided that any jurisdictional authority will accept the Listing found herein. The appropriate authorities should be contacted regarding the acceptability of any given Listing. Visit the QAI Online Listing Directory located at www.qai.org for the most up to date version of this Listing and to validate that this QAI Listing is active. Questions regarding this listing may be directed to info@qai.org. Please include the listing number in the request.