



ELECTRICAL LISTING PROGRAM

Customer: Advance Wire Products Ltd.
 Class: Equipment for the heating of materials
 Location: Surrey, BC
 Listing No. E11393
 Report No. E11393-2401
 Project No. E11393-2401
 Effective Date: Feb 10, 2025
 Last Revised N/A
 Date:
 Expires: <N/A>

Standards: CAN/CSA-C22.2 NO. 61010-1-12 (R2022)/ UL 61010-1: Ed. 3: Safety Requirements for Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements

CSA-C22.2 NO. 61010-2-010:2019 / UL 61010-2-010: Ed. 3: Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use - Part 2-010: Particular Requirements for Laboratory Equipment for the Heating of Materials

Product: Hot Hold Case for indoor

Markings: Products are marked in a permanent manner where it is readily visible after installation with the following:

- a) Manufacturer's name or trademark
- b) Model designation
- c) Month and year of manufacture
- d) Product ratings: voltage, frequency, current
- e) QAI file Number: E11393
- f) The wording "CAN/CSA-C22.2 NO. 61010-1/ UL 61010-1 and CSA-C22.2 NO. 61010-2-010/ UL 61010-2-010"
- g) Warning markings as required by the standards in English
- h) QAI logo shown here:



Models /
Ratings:

Model #	Multiple Listee Model #	Volts	Current	Hz	Location



RAK-0773-B and RAK-0773-F (where B= control panel located at the back and F= control panel located at the front)	N/A	100-120 Vac	10A	60	indoor
RAK-0776	N/A	100-120 Vac	5A	60	indoor