



## ELECTRICAL LISTING PROGRAM

Customer: Cannabix Technologies Inc  
Class: Measurement equipment  
Location: Burnaby

Listing No. E11333-1  
Report E11333-1 Ed. 1 (QAI-6720)  
No.  
Project E11333-1 Ed. 1 (QAI-6720)  
No.  
Effective Apr 30 ,2026  
Date:  
Last N/A  
Revised  
Date:  
Expires: <N/A>

Standards : CAN/CSA-C22.2 NO. 61010-1-12 (R2022)- Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements

UL 61010-1 (Ed.3) - Standard for Safety Requirements for Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements

Product: Measurement equipment  
Markings: Products are marked in a permanent manner where it is readily visible after installation with the following:  
a) Manufacturer's name or trademark  
b) Model designation  
c) Month and year of manufacture  
d) Product ratings: voltage, watts  
e) QAI file Number: E11333-1  
f) QAI logo shown here:



Models / Ratings:	Model #	Multiple Listee Model #	Volts	Wattage	Location
	BCU	N/A	18 Vdc	20 W max	Indoor and Outdoor (IPx2)



The materials, products or systems listed herein have been qualified to bear the QAI Listing Mark under the conditions stated with each Listing. Only those products bearing the QAI Listing Mark are considered to be listed by QAI. No warrantee is expressed or implied, and no guarantee is provided that any jurisdictional authority will accept the Listing found herein. The appropriate authorities should be contacted regarding the acceptability of any given Listing. Visit the QAI Online Listing Directory located at [www.qai.org](http://www.qai.org) for the most up to date version of this Listing and to validate that this QAI Listing is active. Questions regarding this listing may be directed to [info@qai.org](mailto:info@qai.org). Please include the listing number in the request.

\*\*\*