



ELECTRICAL LISTING PROGRAM

Customer: AZCO Industries Ltd.
Class: Laboratory Equipment
Location: Langley, BC, Canada

Listing No. E10552-1
Report No. E10552-1 Ed.1 (QA-6512)
Project No. E10552-1 Ed.1 (QA-6512)
Effective Date: 2026-07-15
Last Revised N/A
Expires: N/A

Standards: CAN/CSA C22.2 No. 61010-1 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements.

UL 61010-1 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 1: General Requirements.

Product: Ozone Generator; Cord-Connected, Laboratory, Dry Locations Only

Markings: Products are marked in a permanent manner where it is readily visible after installation with the following:

- a) Manufacturer's name or trademark
- b) Model designation
- c) Month and year of manufacture
- d) Product ratings: voltage, frequency, watts
- e) QAI file Number: E10552
- f) The wording "CSA CAN/CSA-C22.2 No. 61010-1-12 Ed.3 and UL 61010-1 Ed.3"
- g) Warning markings as required by the standards in English and French
- h) QAI logo shown here:



Models / Ratings:

Model #	Volts	Current	Hz	Location
RMU-DG4	120-240	6.5 A max	50/60	Dry
RMU-K4	120-240	6.5A max	50/60	Dry