



ELECTRICAL LISTING PROGRAM

Customer: Vivitro Labs Inc.
Class: Measurement equipment
Location: Victoria BC
Website: <https://vivitrolabs.com/>

Listing No. E11462-1
Report No. E11462-1, Edition 1
Project No. E11462-1, Edition 1 (QA-7729)
Effective Date: August 27th, 2026
Last Revised August 27th, 2026
Date:
Expires: N/A

Standards: CAN/CSA C22.2 No. 61010-1:12 (R2022) "Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General requirements"

UL 61010-1: Ed. 3 "Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General requirements"

Product: Flow Measurement System

Markings: Products are marked in a permanent manner where it is readily visible after installation with the following:

- a) Manufacturer's name or trademark
- b) Model designation
- c) Month and year of manufacture
- d) Product ratings: voltage, frequency, watts
- e) QAI file Number: E11462-1
- f) The wording "CSA C22.2 No. 61010-1 and UL 61010-1"
- g) Warning markings as required by the standards in English and French
- h) QAI logo shown here:





Models / Ratings:

Model	Rating	Power	Description	Environmental Locations
FMS	100-240Vac	65W _{max}	Flow Measurement System consists of Process Module and Flow Probe	Indoor Dry (Process Module) Indoor Wet (Flow Probe)

The materials, products or systems listed herein have been qualified to bear the QAI Listing Mark under the conditions stated with each Listing. Only those products bearing the QAI Listing Mark are considered to be listed by QAI. No warrantee is expressed or implied, and no guarantee is provided that any jurisdictional authority will accept the Listing found herein. The appropriate authorities should be contacted regarding the acceptability of any given Listing. Visit the QAI Online Listing Directory located at www.qai.org for the most up to date version of this Listing and to validate that this QAI Listing is active. Questions regarding this listing may be directed to info@qai.org. Please include the listing number in the request.
